



## Apogee 3300 Neo

— *Digital Trolley Color  
Doppler System*

### Appearance

- Ergonomic appearance
- Swivel keyboard and monitor
- High resolution color monitor
- 18.5-inch LCD monitor
- Visual Angle:
  - Left and right side:178°
  - Up and down:178°
  - Resolution : 1366×768
- Backlit keyboard, 8 TGC
- Transparent keyboard membrane for minor languages: Spanish, German, Russian, French
- Four active probe connectors
- Six probe holders

### Probe

#### Transducer Types

- Electronic convex probe
- Electronic micro convex probe
- Electronic linear probe
- Electronic trans vaginal probe
- Electronic biplane probe
- Electronic phased array probe
- 4D convex probe
- 4D trans vaginal probe

### Technology

#### Applications

- Abdomen, Urology, Gynecology,
- Obstetrics (1<sup>st</sup> Trimester, 2<sup>nd</sup> and 3<sup>rd</sup> Trimesters), Fetal echo, Multifetation
- Abdomen (PEN), Urology (PEN)
- Thyroid, Breast, Testes, Peripheral vascular, Orthopedics, Podiatry,

- Superficial, Small part (PEN),  
Musculoskeletal Neurology
- Carotid, Vascular (PEN)
- Cardiology, Cardiology (PEN),
- Paediatrics Cardiac

## Highlight

- Auto-Fit: Automatic Optimization
- Nanoview : Speckled Reduction
- Trapezoidal/ Extend Imaging
- Tissue Harmonic imaging
- Edit the exam type and save the user-defined items
- MFI (Inversion THI)
- XBeam : Compound Imaging
- Panoscope : Panoramic Imaging (Optional)
- Auto IMT
- Auto Volume
- Semi-Auto EF
- CFM/CPA/DPA
- VS Flow (Optional)
- Pulse Wave Doppler
- Duplex / Triplex
- HPRF
- CW (Optional)
- Anatomic M mode (3 lines) (Optional)
- Color M mode (Optional)
- TDI (Optional)
- ECG (Optional)
- TEI index
- Free hand 3D
- 4D Lite (Optional)
- 4D Pro (Optional)  
nSlice/AnyCut/QCut
- Elastography (Optional)
- Needle Enhancement(Optional)
- SonoAir : transmit images to iPad/iPhone or the wireless Printer (Optional)
- Smarchive
- DICOM 3.0 (Optional)

## Display mode

- B, 2B, 4B mode
- M, B/M mode
- Color flow mode
- Pulse Wave Doppler
- B/CFM, B/PDI mode
- B/PW mode
- B/CFM/PW, B/PDI/PW
- CW, B/CW, B/CFM/CW mode
- B/E, E mode
- Split B/Color real time mode
- 3D, 4D mode

## Zoom

- HD Zoom:×1.0~×9.0
- Full-View Zoom:×1.0~×8.0
- Full Screen Zoom-The image area fills into whole screen

## Focus

- Continuous dynamic focus
- Dynamic apodization
- Dynamic aperture
- 1~8 selectable transmit focus
- Acoustic lens focus

## Memory

- B-mode cine avi record (60 min)
- Cine-memory
- B-mode (max.2000 frames)
- M-mode (650 s)
- PW-mode (650 s)
- CW-mode (320 s)
- Hard disk 500 GB

## Imaging Processing

### 2D mode

- 8-step TGC slide pots
- Gain: 0~100
- Depth: 1.6~30.8 cm
- Frequency: 5 steps
- Dynamic range adjustable: 30~180dB
- Edge enhancement: 0~3
- Smooth: 0~3
- Nanoview: 0~6
- Persistence: 0~7
- Chroma: 0~8
- Grayscale: 0~23
- Power:  $-\infty \sim 0$ dB, 0~100%
- B steer:  $-20^\circ \sim +20^\circ$
- Line density: 3 steps
- Inversion: left/right, up/down
- Rotate angle:  $0^\circ \sim 270^\circ$
- Gray Scale: 0~23

## M mode

- Gain: 0~100dB
- Sweep speed: 6 steps
- Maps: 0~23
- Chroma: 0~8

## Color mode

- Gain control: 0~100dB
- Pulse repetition frequency: 0.25KHz~12.5KHz
- Wall filter: 50 steps
- Median Filter: 0~3
- Threshold: 0~10
- Color Maps: 0~10
- Smooth: -3~3
- Color persistence: 0~7
- Line density: 2 steps
- Color enhancement: 6~16
- Speed: 0.1cm/s ~ 298.4cm/s
- Color frequency: 4 steps
- Power: 0~100%,  $-\infty$  dB ~ 0 dB
- Baseline: 17 steps
- Steer:  $-20^\circ \sim +20^\circ$
- Priority: 85 steps

- Sampling volume: 1~128

## PW mode

- Gain: 0~100dB
- D map: 0~23
- Frequency: 3 steps
- Chroma: 0~8
- PRFd: 0.25~25KHz
- Basic line: 31 steps
- Wall filter: 50 steps
- Angle:  $-80^\circ \sim +80^\circ$
- Sampling volume: 0.5~40.0mm
- Volume: 0~100%
- D Speed: 1~6
- Smooth: 0~3
- Power:  $-\infty \sim 0$ dB, 0~100%
- Steer:  $-20^\circ \sim +20^\circ$

## CW mode

- Gain: 0~100dB
- Map: 0~23
- Speed: 1~6
- Volume: 0~100%
- Power:  $-\infty \sim 0$ dB, 0~100%
- Smooth: 0~3
- Chroma: 0~8
- Frequency: depend on probes
- WF: 50 steps
- Angle:  $-80^\circ \sim +80^\circ$
- Scale: 1~9

## 4D Lite mode

- 4D map: 31 steps
- Chroma: 0~4
- Rotate angle:  $0^\circ \sim 270^\circ$
- Threshold: 0~100
- Smooth: 0~3
- Brightness: 0~10

## 4D Pro mode

## Data Sheet

- Zoom: 0.2~2.5
- Color: 0~5
- Smooth: 0~5
- Scan Rate: Low, Mid, High
- Opacity Map: 31 steps
- Opacity Pos: 0~255
- nSlice:
  - Planes: 3~39
  - Space: 0.5~20mm
- Q Cut
- Any Cut

### ECG mode

- Gain: 1~8
- Position: 1~10
- Interval: ON/OFF
- ESP: 0~3
- Color: 1~4
- Hide: ON/OFF

## Measurement & Calculation

### Measurement

#### 2D mode (General)

- Distance
- Trace Length
- Ellipse (area)
- Trace (area)
- Angle (general)
- Angle (cross)
- Auto IMT (intima-media thickness)
- Histogram

#### PW mode

- HR (heart rate)
- Velocity
  - PSC (peak systolic velocity)
  - EDV (end diastolic velocity)
  - S/D (systolic/diastolic)
  - RI (resistance index)

- PG (pressure)
- ACC (acceleration)
- Time
- Manual Trace
  - PSC (peak systolic velocity)
  - EDV (end diastolic velocity)
  - MN (median)
  - ACC (acceleration)
  - S/D (systolic/diastolic)
  - RI (resistance index)
  - PI (pulsatility index)
  - HR (heart rate)
  - PG (pressure)

- Auto Trace
  - PSC (peak systolic velocity)
  - EDV (end diastolic velocity)
  - MN (median)
  - ACC (acceleration)
  - S/D (systolic/diastolic)
  - RI (resistance index)
  - PI (pulsatility index)
  - HR (heart rate)
  - PG (pressure)

- Range Trace
  - PSC (peak systolic velocity)
  - EDV (end diastolic velocity)
  - MN (median)
  - ACC (acceleration)
  - S/D (systolic/diastolic)
  - RI (resistance index)
  - PI (pulsatility index)
  - HR (heart rate)
  - PG (pressure)

### Calculation

#### Abdomen

#### Urology

#### Gynecology

#### Obstetrics (1<sup>st</sup> Trimester)

#### Obstetrics (2<sup>nd</sup> and 3<sup>rd</sup> Trimesters)

#### Fetal echo

Thyroid

Breast

Testes

Neonate

Peripheral vascular

Orthopedics

Carotid

Cardiology

## Physical Features

### Connectivity

- Video out port
- S-Video out port
- Audio in port
- Audio out port
- MIC
- VGA out port
- 3-4 USB port
- Printer control port
- AC power input port
- HDMI digital port
- Network interface
- Foot SW
- ECG port

### Dimension

- Gross dimension
  - Whole set  
950 mm (L) X 670 mm (W) X  
1200 mm (H)
- Net dimension
  - 825 mm (D) X 580 mm (W) X  
1280~1410 mm (H)

### Weight

- Gross weight
  - 90kg
- Net weight
  - 57kg

### Power Requirements

- Voltage  
100-240V±22V~(230±23V for EU countries)
- Frequency  
50Hz±1Hz; 60Hz±1Hz
- Rated Power  
500VA

### Operation Conditions

- Ambient temperature  
0°C to +40°C
- Relative humidity  
30% to 85%
- Atmospheric Pressure  
700hPa to 1060hPa

### Stored Conditions

- Ambient temperature  
-20°C to +60°C
- Relative humidity  
15% to 93%
- Atmospheric Pressure  
500hPa ~ 1060hPa

## Software & Accessories

### Standard Accessories

- Power Cable
- Operation Manual
- Potential equalization conductor
- Printer control cable
- S-Video cable
- Fuse
- BNC/RCA cable
- Dust-proof cover
- Recovery system CD
- Wireless network antenna(if purchase WiFi function)

- 4D training CD

### Optional Accessories

- B/W or color Video printer
- LaserJet or inkjet printer
- Biopsy guide for convex or linear probe
- Biopsy guide for trans vaginal or transrecta probe
- Foot switch
- ECG cable
- Gel warmer

### Applied Standards

#### Quality Standards

- ISO 9001:2008
- ISO 13485:2003

#### Conformance Standards

- UL 60601-1

- EN 60601-1 and IEC 60601-1
- EN 60601-1-1 and IEC 60601-1-1
- EN 60601-1-2 and IEC 60601-1-2
- EN 60601-1-4 and IEC 60601-1-4
- EN 60601-1-6 and IEC 60601-1-6
- EN 60601-2-37 and IEC 60601-2-37
- EN 62304 and IEC 62304

### CE Declaration

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective products / product categories according to Annex II section 3 of the Directive 93/42/EEC on Medical Devices.